



# EC Certificate

## Full Quality Assurance System

Certificate No.:  
**247516-2017-AQ-NOR-NA-PS Rev. 0.0**

Project No.:  
**PRJC-437446-2013-MSL-NOR**

Valid Until:  
**18 September 2018**

This is to certify that the quality system of:

### **Elos Medtech AB**

Bäckedalsvägen 5  
531 3116 Timmersdala  
Sweden

For design, production and final product inspection/testing of:

### **Sterile and Non-Sterile Implants and Instruments**

Has been assessed with respect to:

**The conformity assessment procedure described in Article 11.3.a and Annex II excluding section 4 (Module H2) of Council Directive 93/42/EEC on Medical Devices, as amended**

and found to comply.

Further details of the product(s) and conditions for certification are given overleaf.

Place and Date:  
**Høvik, 15 November 2017**



For:  
**DNV GL NEMKO PRESAFE AS**

**Alessandra Rinna**

The Certificate has been digitally signed.  
See [www.presafe.com/digital\\_signatures](http://www.presafe.com/digital_signatures) for more info

Notice: The Certificate is subject to terms and conditions as set out in the Certification Agreement. Failure to comply may render this Certificate invalid.

# EC Certificate

## Full Quality Assurance System

Certificate No.:  
**247516-2017-AQ-NOR-NA-PS Rev. 0.0**

Project No.:  
**PRJC-437446-2013-MSL-NOR**

Valid Until:  
**18 September 2018**

### Jurisdiction

Application of Council Directive 93/42/EEC of 14 June 1993, adopted as “Forskrift om Medisinsk Utstyr” by the Norwegian Ministry of Health and Care Services.

### Certificate history:

Revision	Description	Issue Date
0.0	Replaces certificate 118365-2012-CE-SWE-NA Rev. 4.0 (NB 0434) following transfer of Notified Body functions to DNV GL Nemko Presafe AS (NB 2460). Removal of devices.	2017-11-15

### Products covered by this Certificate:

Product Description	Product Name	Class
Cresco API™ Kit	Different articles to manufacture a customized implant prosthetic construction connected to different implant platforms.	IIb
Cresco™ Prosthetic Instruments	Instruments	IIa
Retractor	Instruments	IIa

The complete list of devices is filed with the Notified Body

### Sites covered by this certificate

Elos Medtech Timmersdala AB  
Bäckedalsvägen 5, 540 16, Timmersdala, Sweden

# EC Certificate

## Full Quality Assurance System

Certificate No.:  
**247516-2017-AQ-NOR-NA-PS Rev. 0.0**

Project No.:  
**PRJC-437446-2013-MSL-NOR**

Valid Until:  
**18 September 2018**

### Terms and conditions

The certificate is subject to the following terms and conditions:

- Any producer (see 2001/95/EC for a precise definition) is liable for damage caused by a defect in his product(s), in accordance with directive 85/374/EEC, as amended, concerning liability of defective products.
- The certificate is only valid for the products and/or manufacturing premises listed above.
- The Manufacturer shall fulfil the obligations arising out of the quality system as approved and uphold it so that it remains adequate and efficient.
- The Manufacturer shall inform Presafe of any intended updating of the quality system and Presafe will assess the changes and decide if the certificate remains valid.
- Periodical audits will be held, in order to verify that the Manufacturer maintains and applies the quality system. Presafe reserves the right, on a spot basis or based on suspicion, to pay unannounced visits.

The following may render this Certificate invalid:

- Changes in the quality system affecting production.
- Periodical audits not held within the allowed time window.

### Conformity declaration and marking of product

When meeting with the terms and conditions above, the producer may draw up an EC declaration of conformity and legally affix the CE mark followed by the Notified Body identification number of Presafe.

End of Certificate